

# X-Ray Interlock Check Records

|                                 |      |
|---------------------------------|------|
| Building and Room               |      |
| Health Physics Inventory Number | MUX- |
| Facility Manager                |      |

Interlock or safety systems tests will be conducted according to procedures from the manufacturer or facility manager. These checks shall be performed quarterly, except for emergency stop buttons which must be tested annually.

| Date | Pass/Fail | Comments |
|------|-----------|----------|
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |
|      |           |          |

If any of interlock tests fail, cease all use of the x-ray machine and contact Health Physics. Do not use the device without Health Physics approval.

**Health Physics Contact Information:**

Phone: 905-525-9140 x24226  
 Email: [hphys@mcmaster.ca](mailto:hphys@mcmaster.ca)